



405 E Gude Dr, STE 214
Rockville, MD 20850,
USA www.abpbio.com

CERTIFICATE OF ANALYSIS

Catalog Number N023
Product Name iQuant™ microRNA Assay Kit
Lot Number AB2303N1

	LOT DATA	SPECIFICATION
Absorption Optical Density of Component D	0.232 at 260 nm	0.23 ± 0.008 at 260 nm
Limit of Detection Sensitivity	0.5 ng/well	≤ 0.5 ng/well
Linear Regression R-squared	0.998	≥ 0.98